

Jakarta, Feb 2017

Dear Patient

Jakarta

Herewith we send usage data downloaded APAP (S9 autoSet Spirit), according to data obtained from the amount of (Jan – Dec, 2017) can be explained summary as follows:

1. Setting Mode Therapy: AutoSet (4.0 cmH₂O - 15.0 cmH₂O)
2. AHI Report:
 - AHI (Apnea Hypopnea Index) dated Jan to 7 Feb 2017: 1.0 (Normal)

Range condition of OSA:

- 0 - <5 events per hour ◊ Normal
- > 5 - <15 events per hour ◊ Light (Mild)
- > 15 - <30 events per hour ◊ Medium (Moderate)
- > 30 events per hour ◊ Weight (Severe)

3. The maximum mask leakage under normal threshold is 14.4 l / min (normal 24 l / min)
4. The use of the APAP pressure an average of 6.5 cm H₂O.
5. Please ensure good conditions for use of the mask in order to minimize leakage.

Thus the results of APAP downloads you have, for your attention and cooperation we would like to thank.

Best regards,

Resindo Medika